

上海市计量测试技术研究院 华东国家计量测试中心

检定证书

Verification Certificate

证书编号: 2015K22-20-501216 号
Certificate No.

送检单位
Applicant

上海交通大学医学院附属新华医院 (6号楼7楼)

计量器具名称
Name of Instrument

体重秤

型号/规格
Type/ Specification

RGZ-120/120kg e=500g

出厂编号
Series No.

094026

制造单位
Manufacturer

上海东方衡器有限公司

检定依据
Verification Regulation

JJG13-1997 模拟指示秤检定规程

检定结论
Conclusion

III级 合格

(检定专用章)
Stamp

批准人
Approved by

核验员
Checked by

检定员
Verified by

检定日期 2015 年 08 月 25 日
Date for Verification Year Month Day

有效期至 2016 年 08 月 24 日
Valid until Year Month Day

计量检定机构授权证书号: (国)法计(2012)01019号/01039号
Authorization Certificate No.

电话: 021-38839800
Telephone

地址: 上海市张衡路 1500 号(总部)
Address No.1500 Zhangheng Road, Shanghai (headquarters)

邮编: 201203
Post Code

传真: 021-50798390
Fax

网址: www.simt.com.cn
Web site



证书编号: 2015K22-20-501216
Certificate No.

SHANGHAI INSTITUTE OF MEASUREMENT AND TESTING TECHNOLOGY
NATIONAL CENTER OF MEASUREMENT AND TEST FOR EAST CHINA

本次检定所使用的计量(基)标准/主要计量器具:
Measurement standards / Measuring instrument used in this verification

名称/型号 Name/Model	编号 Number	证书编号/有效期限 Certificate No./Due date	测量范围/准确度等级或 最大允差或不确定度 Measurement range/accuracy class or maximum permissible errors or uncertainty of measurement
砝码/50mg~5kg	A ₂ -21	2015K21-20-800126/2016.06.12	50mg~11.11105kg/ M ₁ 等级
砝码/25kg	B ₂ -1-138	2015K21-20-800128/2016.06.12	(25~3450) kg /M ₁ 等级

以上计量标准器具的量值溯源至国家基准。
The value of a quantity of measurement standard used in this verification is traced to those of the national primary standards in the P.R. China.

检定地点及环境条件:
Location and environmental condition for the verification

地点: 控江路 1665 号
Location

温度: 22 °C; 湿度: / %RH; 其它: /
Ambient temperature Humidity Others

检定结果/说明:
Results of verification and additional explanation

检定项目		技术要求	检定结果
外观检查		铭牌信息应完整	符合要求
空秤变动性		$\pm 0.5e$	0.0e
偏载测试		$\pm 1.0e$	0.3e
称量 测试	0~50e	$\pm 0.5e$	0.2e
	>50e~200e	$\pm 1.0e$	0.4e
	>200e~Max	$\pm 1.5e$	0.5e
灵敏度		加上 1e 砝码,指针位移不小于 7/10 分度	符合要求
重复性		$\leq 1.0e$	0.2e

检定结果内容结束

本证书提供的结果仅对本次被检的器具有效。未经本院/中心批准, 部分采用本证书内容无效。
The data are valid only for the instrument(s). Partly using this certificate will not be admitted unless allowed by SIMT.